

Patient Alert PA- 5 FLOOR

Fall Prevention System QUICK START GUIDE

Emergency Caller Products 

Emergency Caller Products (T) 800-227-2474 (F) 800-762-7613
www.emergencycaller.com sales@emergencycaller.com
U.S.A. and World Patents Issued and Pending on Certain Products



Follow these simple instructions to set up your system

Step 1. Positioning the Pad.

Place the sensor pad next to bed, in front of chair, or area to detect patient standing or walking, with the cord placed unobtrusive and routed to Patient Alert unit.



Step 2. Connect the cord.

Place the connector plug on the end of the pad into the socket on top of the Patient Alert unit as shown.



Step 3. Position the Alert Unit.

With the Patient Alert switch in the **off** position, attach the Patient Alert unit to desired location on bed or chair using the Mesh Bag provided. Patient Alert sound is designed to alert care-taker while alerting patient to return to bed or chair or proper location.

Step 4. Operate the system.

Place the patient in position, with patient weight **OFF** the Floor Pad. Now you are ready to activate the Patient Alert unit.

***IMPORTANT**

**If Patient Alert System Fails -
Replace Battery.
Replace Battery Periodically**

Test System Before Each Use

A- Momentary Activation (MOM)

Switch Alert unit to **MOM** position. This will activate Alarm when patient steps on or weight is on floor pad. The Patient Alert unit will Automatically reset and silence alarm when patient weight is Off Sensor Pad or unit is turned off by caretaker.

B- Continuous Activation (CONT)

When the Alert switch is placed in **CONT** position and patient steps on or weight is on Sensor Pad, Patient Alert will sound alarm continuously until turned off by caretaker



**LIFETIME SENSOR PAD INSTALLATION INSTRUCTIONS
PLEASE READ CAREFULLY****FLOOR SENSOR PADS
CHAIR, BED or AREA****FLOOR SENSOR PADS
BEDSIDE , AREA USE****Step 1. Positioning the Sensor Pad**

Uncoil the cord from the sensor pad. Place the sensor pad on the floor with the cord coming straight under towards the back of the bed towards the alarm unit. The pad would normally be placed on the floor below waist level and the alarm would sound when the patient places weight on the pad.

The RJ22 connection at the end of the cord may be connected into any of Emergency Caller Products alarm units such as the Patient Alert (PA), the Extended Range Transmitter, TR-300, or the Long Range Transmitter, the LT-900. Consult your Emergency Caller Products sales representative to learn how these other products work with this model sensor pad. The pad may also be connected into other products that will receive the RJ22 connector.

Sensor Pads detect both Applied Weight "Patient ON" pad or Removed Weight "Patient Off". Selection made by switch on Universal Transmitter Unit or other monitoring products. Can be used to detect patient in bathtub or shower for entry or exit (weight off or weight on). Sensor pads are slip resistant and completely waterproof.

Step 2. Chair - Bed Use

Place the sensor pad next to bed, in front of chair, or area to detect patient standing or walking, with the cord placed unobtrusive and routed to Patient Alert unit or other Sensor device.

DO NOT FOLD, BEND OR ROLL SENSOR PAD

* (This will Void Lifetime Warranty)

IMPORTANT WARRANTY INFORMATION**Do Not Return Defective Sensor Pads to Distributor!****If you believe Sensor Pad is defective, Call 1-800-227-2474****To speak with Customer Service Tech and receive Return Authorization Number**